

Customer Information

Customer:CUSTOMER NAME
CUSTOMER ADDRESS
CITY, STATE ZIP CODE

Purchase Order : 8400

Equipment Information

Asset :	00075	Serial Number:	6111-326/981198/1054
Description:	HIPOT TESTER	Model Number:	MHP2
Manufacturer:	AVO MEGGER	Cal. Interval:	12 MONTHS
Accuracy:	Manufacturer Specifications		

Event Information

Service Requested:	CALIBRATION	Cal Date:	11/10/2021
Condition Received:	IN TOLERANCE	Cal Due Date:	11/10/2022
Condition Returned:	LEFT AS FOUND	Temp./RH:	23 C / 34 %
Cal Procedure:	33K1-4-2006-1	AC/DC HYPOT TESTERS	
		Technician:	L WILCOX
Calibration Notes:		QA Approval:	Craig Fink, QA Representative

This document certifies that the unit conformed to applicable specifications upon successful completion of the calibration. Any number of factors may cause the calibration item to drift out of calibration before the recommended calibration interval has expired.

This instrument has been calibrated using test equipment of a known accuracy, traceable to the National Institute Of Standards and Technology (NIST) or other recognized metrology institutes. Trident Calibration Labs' policies and procedures comply with ANSI/NC SL Z540.1-1994 as well as the ISO 9000 Series Quality Standards.

A test uncertainty ratio (TUR) of 4:1 (k=2, approx.. 95% confidence level) is maintained unless otherwise stated.

This certificate shall not be reproduced, except in full, without written consent of Trident Calibration Labs.

Standards

<u>I.D.</u>	<u>Mfg</u>	<u>Model #</u>	<u>Description</u>	<u>Cal. Due Date</u>	<u>NIST Traceability #</u>
10-0079	GENERAL RADIO	1433-M	DECADE RESISTOR	7/5/2022	159857
10-0426	FLUKE	8508A	DIGITAL MULTIMETER	11/1/2022	169744
10-0444	VITREK	4700	PRECISION HV METER	7/22/2022	939D0B66